

BIOSENSORS: PRINCIPLES, TYPES, NANOBIOSENSOR TECHNOLOGIES, AND BIOMEDICAL APPLICATIONS—A COMPREHENSIVE REVIEW

Khushbu Rani Sahu¹, Ankit Jain²

¹Associate Professor, Jai Narain College of Pharmacy, JNCT Professional University,
Bhopal, Madhya Pradesh, India

²Professor, Shri Rawatpura Sarkar Pharmacy College, Sagar, Madhya Pradesh, India

Abstract

Biosensors are analytical devices that combine a biological recognition element with a physicochemical transducer to detect and quantify specific analytes. Their selectivity, sensitivity, rapid response, portability, and potential for low-cost operation have made them important tools in biomedical analysis, environmental monitoring, food analysis, and pharmaceutical applications. This review summarizes the fundamental principles of biosensing, major recognition and transduction components, important performance characteristics, and the development requirements of practical biosensors. It further discusses electrochemical and optical biosensors, with emphasis on nanobiosensors and their role in improving analytical sensitivity through nanomaterials. Applications in glucose monitoring, metabolic assessment, environmental detection, and biomedical diagnostics are also reviewed. The article highlights the continuing development of miniaturized, portable, and sensitive biosensing systems for clinical and other analytical applications.

Keywords

Biosensors; Bioreceptors; Transducers; Electrochemical biosensors; Optical biosensors; Nanobiosensors; Glucose monitoring; Biomedical applications

1. Introduction

Biosensors are analytical measuring devices that incorporate a biological recognition element and a transducer. They combine the selective recognition properties of biological systems with the processing capabilities of modern microelectronics and optoelectronics. Consequently, biosensors have important applications in medicine, environmental monitoring, food analysis, and pharmaceutical processing [1,2].

The biological component may include enzymes, antibodies, nucleic acids, cells, tissues, organelles, or other recognition materials that interact selectively with a target analyte. The physicochemical component, generally called the transducer, converts the interaction between the analyte and the recognition element into a measurable signal. Depending on the transduction mechanism, the output may be potentiometric, amperometric, conductometric, optical, calorimetric, gravimetric, or otherwise physicochemical in nature [1-4].

For effective operation, the analyte must reach the recognition layer, where a specific biochemical interaction produces a measurable change. The transducer subsequently converts this change into an analytical signal related to analyte concentration. The combination of biological selectivity and physical signal conversion is the fundamental basis of biosensor technology.

2. Fundamental Principle and Components of Biosensors

The basic biosensor consists of three principal components: a bioreceptor, a transducer, and a signal-processing/readout system. The bioreceptor is an immobilized biological recognition element that selectively interacts with the analyte. Common bioreceptors include enzymes, antibodies, DNA or other nucleic-acid probes, whole microbial or mammalian cells, tissues, organelles, lectins, and synthetic recognition materials [1,3].

The transducer converts the biochemical interaction into an electrical, optical, thermal, mechanical, or other measurable signal. The magnitude of the generated signal is generally related directly or indirectly to the concentration of the analyte. Electrochemical transducers are particularly important because they can provide relatively simple, compact, and low-cost analytical systems [3,4].

Biosensors may therefore be classified according to both the biological recognition element and the transduction principle. On the basis of transduction, major groups include electrochemical, optical, piezoelectric, thermal, and related sensor systems. Electrochemical biosensors can be further categorized as potentiometric, amperometric, conductometric, impedimetric, and field-effect devices [3,4].

3. Essential Performance Characteristics

Selectivity is a fundamental characteristic of a biosensor because the device should respond preferentially to the target analyte in the presence of other substances. Antigen-antibody interactions are examples of highly selective biological recognition. Accuracy and precision are important for quantitative measurements, while signal stability is essential for continuous or repeated monitoring.

Sensitivity refers to the ability of a sensor to detect small changes or low concentrations of an analyte. The working range is the concentration interval over which the sensor provides an acceptable analytical response. Response time is the time required to obtain a measurable response after exposure to the analyte, whereas recovery time is the time required for the sensor to return to its operating condition. Operational lifetime and the number of usable measurement cycles are also important because biological recognition components may gradually lose activity. In some commercial devices, disposable transducers or sensor elements are replaced after each measurement [5].

4. Development Requirements for Biosensors

Successful biosensor development requires a biological recognition element with high specificity, adequate stability, and low variation between measurements. Ideally, the sensing reaction should be relatively independent of physical parameters such as mixing, pH, and temperature so that samples can be analyzed with minimal pretreatment.

When cofactors or coenzymes are required, their co-immobilization with the biological component can facilitate sensor operation. The analytical response should be accurate, precise, reproducible, and sufficiently linear over the concentration range of interest. Electrical and transducer-generated noise should be minimized. For invasive clinical

monitoring, the sensing system should be small, biocompatible, and free from undesirable toxic or antigenic effects. Practical systems should also resist inactivation and degradation and, where rapid clinical measurements are required, should provide near-real-time analysis. Low cost, portability, ease of use, and suitability for operators with limited technical training are desirable characteristics [6].

5. Types of Biosensors

Biosensors can be categorized according to their transduction mechanisms. Important groups include electrochemical, optical, piezoelectric, thermal, and other physicochemical sensor systems. Among these, electrochemical and optical biosensors have been extensively investigated because of their sensitivity, adaptability, and potential for miniaturization [3,6].

5.1 Electrochemical Biosensors

Electrochemical biosensors measure electrical changes generated by biochemical interactions. Amperometric devices measure current, potentiometric devices measure potential or charge-related changes, and conductometric devices measure changes in conductivity. Impedimetric systems determine changes in impedance, while field-effect devices use transistor-based configurations to detect electrical responses associated with biochemical interactions [3,6].

Because many electrochemical reactions occur near the electrode surface, electrode design and material strongly influence sensor performance. A typical electrochemical configuration may include a reference electrode, a counter electrode, and a working electrode. The reference electrode, commonly based on Ag/AgCl, provides a stable reference potential. The working electrode participates in the sensing process, while the counter electrode completes the electrical circuit. Platinum, gold, carbon-based materials, and related conductive materials are frequently used depending on the application [6].

Nanotechnology has expanded the design possibilities of electrochemical biosensors. Nanoparticles, nanowires, carbon nanotubes, and nanorods provide high surface-area-to-volume ratios and can facilitate efficient immobilization of recognition elements. Their dimensions are comparable to many biomolecular structures, which may improve

analyte capture and analytical sensitivity. Nanostructured materials have consequently become important components in DNA sensing, pathogen detection, and other bioelectronic applications [17,18,24].

5.2 Nanobiosensors and Nanomaterials

Nanobiosensors are biosensing systems that incorporate nanomaterials or nanostructures. Nanomaterials generally possess at least one dimension in the nanoscale range and may exhibit physicochemical properties that differ from their bulk counterparts. Nanotubes, nanowires, nanorods, nanoparticles, and nanocrystalline thin films have been investigated as components of sensing and transduction systems [17,18].

The large surface area of nanomaterials can increase the available sites for immobilization of biological recognition elements and may improve analyte capture. Their electrical and mechanical properties can also support sensitive transduction. Integration of nanomaterials with electrical systems has contributed to the development of miniaturized nanoelectromechanical and bioelectronic devices [17,18].

A nanobiosensor generally includes a target analyte, a bioreceptor, a transducer, and a detection/readout system. Immobilization of the bioreceptor is an important step because it places the recognition element in a stable configuration for interaction with the analyte. Sensor performance can nevertheless be affected by temperature, pH, interfering substances, and other physicochemical variables, which must be considered during design and validation [17,18].

5.3 Optical Biosensors and Surface Plasmon Resonance

Optical biosensors detect changes in optical properties produced by biochemical interactions. Surface plasmon resonance (SPR) is an important optical sensing principle in which biomolecular binding at a metal surface changes the optical response of the sensor. A typical SPR sensor contains a thin gold layer associated with a glass substrate and an optical detection arrangement. The surface can be modified to facilitate immobilization of molecules that selectively bind the target analyte [8,10,11].

When biomolecules bind to the functionalized surface, the local refractive index changes and the SPR response shifts. This change can be measured with high sensitivity and

used to monitor molecular interactions. Other optical biosensors rely on changes in absorbance or fluorescence and may be designed for specific analytical applications [8,11].

6. Biomedical and Clinical Applications

Biosensors have important applications in healthcare because rapid measurement of biochemical parameters can support clinical decision-making and monitoring. Measurements of blood chemistry are particularly relevant in intensive care and other settings in which biochemical changes can occur rapidly. Conventional laboratory analysis often requires collection and transport of discrete samples, whereas biosensors have the potential to provide rapid or continuous measurements at the point of care [15,20].

6.1 Glucose Monitoring

Blood glucose monitoring is one of the most established applications of biosensor technology. A conventional enzymatic glucose biosensor uses glucose oxidase as the biological recognition element. Glucose is oxidized through an enzyme-mediated process, and the resulting electrochemical changes are measured at the electrode. The current generated is related to glucose concentration [15,20].

Glucose biosensors have enabled portable monitoring systems that can be used by patients and healthcare professionals. Developments in sensor design have also supported investigation of continuous glucose monitoring and closed-loop insulin delivery systems. Such systems use information from glucose sensors to support more responsive insulin administration, although sensor performance, calibration, biocompatibility, and long-term stability remain important considerations [16,25].

6.2 Insulin Therapy and Closed-Loop Systems

Improved management of insulin-dependent diabetes has encouraged the development of continuous insulin-delivery systems. Regardless of the delivery method, insulin administration must be informed by current blood glucose levels. Traditional approaches rely on intermittent glucose measurements, including finger-prick testing, whereas closed-loop systems integrate glucose sensing with insulin delivery.

In a closed-loop configuration, the output of a glucose sensor provides information to an insulin-delivery device. The overall objective is to respond more appropriately to changes in glucose concentration while reducing the limitations associated with intermittent measurement and manual adjustment. Biosensor research therefore remains relevant to the development of integrated systems for diabetes management [15,16,25].

7. Environmental and Other Applications

Biosensors are also applicable to environmental monitoring, food analysis, and detection of hazardous substances. Biological recognition systems can detect target compounds at concentrations that may be difficult to identify by conventional approaches. Applications include monitoring of pollutants, trace gases, water quality, and other environmental contaminants [11].

Optical and nanomaterial-based biosensors have also been explored for food analysis and the detection of specific biological or chemical components. Their selectivity, rapid response, and potential for miniaturization make them attractive for applications in which rapid analytical information is required [8,11,17].

8. Market Potential and Future Prospects

The continuing demand for inexpensive, reliable, portable, and rapid analytical devices supports the development of biosensors for hospitals, emergency departments, operating rooms, laboratories, and home-based monitoring. Point-of-care testing is an important area because rapid measurements can reduce dependence on centralized laboratory facilities and may provide information close to the time of clinical decision-making.

Future development is likely to emphasize miniaturization, nanostructured materials, improved immobilization strategies, multiplex detection, continuous monitoring, and integration with digital readout systems. The practical success of biosensors will depend on maintaining selectivity and sensitivity while improving stability, reproducibility, affordability, and ease of use [17,18,24].

9. Conclusion

Biosensors integrate selective biological recognition with physicochemical signal transduction and have become important analytical tools in biomedical, environmental, food, and pharmaceutical applications. Electrochemical and optical approaches provide versatile platforms, while nanomaterials offer additional opportunities for improving surface area, recognition-element immobilization, and signal transduction. Glucose monitoring demonstrates the clinical value of biosensors, whereas environmental and food applications illustrate their broader analytical utility. Continued advances in miniaturization, nanotechnology, immobilization, and integrated detection systems are expected to support the development of increasingly sensitive, rapid, portable, and practical biosensing devices.

References

1. Khandpur RS. Handbook of biomedical instrumentation. New Delhi: Tata McGraw-Hill Education; 1992. p. 15-75.
2. Buerk DG. Biosensors: theory and applications. Boca Raton: CRC Press; 1995.
3. Newman JD, Setford SJ. Enzymatic biosensors. *Mol Biotechnol.* 2006;32(3):249-268.
4. Eggins BR. Chemical sensors and biosensors. Chichester: John Wiley & Sons; 2008.
5. Grieshaber D, MacKenzie R, Vörös J, Reimhult E. Electrochemical biosensors—sensor principles and architectures. *Sensors.* 2008;8(3):1400-1458.
6. Kumar A. Biosensors based on piezoelectric crystal detectors: theory and application. *JOM-e.* 2000;52(10).
7. Martins TD, Ribeiro ACC, de Camargo HS, da Costa Filho PA, Cavalcante HPM, Dias DL. New insights on optical biosensors: techniques, construction and application. 2013.
8. Strehlitz B, Nikolaus N, Stoltenburg R. Protein detection with aptamer biosensors. *Sensors.* 2008;8(7):4296-4307.
9. Pohanka M, Skládal P. Electrochemical biosensors—principles and applications. *J Appl Biomed.* 2008;6(2):57-64.
10. da Costa Silva LM, Melo AF, Salgado AM. Biosensors for environmental applications.

11. Allain LR, Stratis-Cullum DN, Vo-Dinh T. Investigation of microfabrication of biological sample arrays using piezoelectric and bubble-jet printing technologies. *Anal Chim Acta*. 2004;518(1):77-85.
12. Yoo EH, Lee SY. Glucose biosensors: an overview of use in clinical practice. *Sensors (Basel)*. 2010;10(5):4558-4576.
13. Blum LJ, Coulet PR. *Biosensor principles and applications*. 1991.
14. Croce RA Jr, Vaddiraju S, Papadimitrakopoulos F, Jain FC. Theoretical analysis of the performance of glucose sensors with layer-by-layer assembled outer membranes. *Sensors (Basel)*. 2012;12(10):13402-13416.
15. Malik P, Katyal V, Malik V, Asatkar A, Inwati G, Mukherjee TK. Nanobiosensors: concepts and variations. *ISRN Nanomater*. 2013;2013:327435. 9 p.
16. Rai M, Gade A, Gaikwad S, Marcato PD, Durán N. Biomedical applications of nanobiosensors: the state-of-the-art. *J Braz Chem Soc*. 2012;23(1):14-24.
17. Shtenberg G, Segal E. *Porous silicon optical biosensors*. 2014.
18. Chambers JP, Arulanandam BP, Matta LL, Weis A, Valdes JJ. *Biosensor recognition elements*. San Antonio: Texas University at San Antonio, Department of Biology; 2008.
19. Kawai T, Akira S. The role of pattern-recognition receptors in innate immunity: update on Toll-like receptors. *Nat Immunol*. 2010;11(5):373-384.
20. Zhang X, Ju H, Wang J, editors. *Electrochemical sensors, biosensors and their biomedical applications*. Amsterdam: Academic Press; 2011.
21. Valgimigli F, Mastrantonio F, Lucarelli F. Blood glucose monitoring systems. In: *Security and privacy for implantable medical devices*. New York: Springer; 2014. p. 15-82.
22. Heise HM, Marbach R, Koschinsky T, Gries FA. Multicomponent assay for blood substrates in human plasma by mid-infrared spectroscopy and its evaluation for clinical analysis. *Appl Spectrosc*. 1994;48(1):85-95.

23. Park S, Boo H, Chung TD. Electrochemical non-enzymatic glucose sensors. *Anal Chim Acta*. 2006;556(1):46-57.
24. Vasylieva N. Implantable microelectrode biosensors for neurochemical monitoring of brain functioning [doctoral dissertation]. Zurich: ETH Zurich; 2013.
25. Henry C. Getting under the skin: implantable glucose sensors. *Anal Chem*. 1998;70(17):594A-598A.
26. Abad JM, Vélez M, Santamaría C, Guisán JM, Matheus PR, Vázquez L, et al. Immobilization of peroxidase glycoprotein on gold electrodes modified with mixed epoxy-boronic acid monolayers. *J Am Chem Soc*. 2002;124(43):12845-12853.
27. Huang S. Glucose biosensor using electrospun Mn₂O₃-Ag nanofibers. 2011.
28. Toghill KE, Compton RG. Electrochemical non-enzymatic glucose sensors: a perspective and an evaluation. *Int J Electrochem Sci*. 2010;5(9):1246-1301.

How to Cite This Article

Sahu KR, Jain A. Biosensors: Principles, Types, Nanobiosensor Technologies, and Biomedical Applications—A Comprehensive Review, *Panacea Journal of Pharmacy and Pharmaceutical Sciences* 2026:15(3), 38-47.